

# Certificate of Conformity

## NucleoCounter® NC-100™

**Name of product:** NucleoCounter® NC-100™

**Type:** Mammalian Cells

**Other identifying data:** Part no. 900-0004

**The product complies with requirements of the following directives:**

- 89/336/EEC - Electromagnetic Compatibility (EMC)
- 89/392/EEC - Machinery Directive

**Harmonized standards, which have been used:**

EN61326: 1997+A1: 1998 (Class B) +A2: 2000, Electrical equipment for measurement, control and laboratory use – EMC requirements (emission and immunity). Annex B and C from A1: 1998 is used.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Date: 2013-05-30

Signed:



Name: Sarah Rie Rasmussen  
Position: QA/QC Manager

**Name and address of manufacturer:**

ChemoMetec A/S  
Gydevang 43  
DK-3450 Allerød  
DENMARK



chemometec

ChemoMetec A/S  
Gydevang 43  
DK-3450 Allerød  
Denmark

Phone(+45) 48 13 10 20  
Fax (+45) 48 13 10 21  
Mail [info@chemometec.com](mailto:info@chemometec.com)  
Web [www.chemometec.com](http://www.chemometec.com)