

Certificate of Conformity

NucleoCounter® SP-100™

Name of product: NucleoCounter® SP-100™

Type: Mammalian Semen

Other identifying data: Part no. 900-0100

The product complies with requirements of the following directives:

- 89/336/EEC - Electromagnetic Compatibility (EMC)
- 89/392/EEC - Machinery Directive

Harmonized standards, which have been used:

EN61326: 1997+A1: 1998 (Class B) +A2: 2000, Electrical equipment for measurement, control and laboratory use – EMC requirements (emission and immunity). Annex B and C from A1: 1998 is used.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Date: 2013-05-30

Signed:



Name: Sarah Rie Rasmussen
Position: QA/QC Manager

Name and address of manufacturer:

ChemoMetec A/S
Gydevang 43
DK-3450 Allerød
DENMARK



chemometec

ChemoMetec A/S
Gydevang 43
DK-3450 Allerød
Denmark

Phone(+45) 48 13 10 20
Fax (+45) 48 13 10 21
Mail info@chemometec.com
Web www.chemometec.com